No.:163

Date: 01/08/2020



## To/ Companies and suppliers

## Sub/ Tendering Announcement

We request your attendance to participate in the tender for provision and installation of fifty-seven (57) Bed Head Units for Heevie project of "Renovation of East Erbil Emergency Hospital" by submitting applications under the conditions listed below. Applications will be received at the organization headquarter in Duhok, St. Mahabad No.7, KRO neighborhood near Sheelan private Hospital. The deadline for receipt of the applications will be on Monday (17/08/2020) at 12:00 pm.

For a copy of the tender and for more information, please call 07504058155

With regards

Heevie Health NGO
محفراوی فیل با ما له می سرسمبر
ظمه هیفی الد سدیه سما در مرسا

Manager

Bakshan Ali Aziz

## **Terms of the Tender:**

Heevie organization wishes to purchase the items involved within the tender according to the following regulations:

- 1- The Company/Supplier must have a valid bank account in the name of the Company/Supplier or the legally authorized representative of the Company/Supplier (the latter should be confirmed by an official letter from the Company/Supplier)
- 2- The tender papers to be presented inside a sealed envelope stamped by the Company/Supplier, provided that it is presented in a single envelope bearing the number of tender.
- 3- The provision and installation of the materials involved within the tender must be in line with the below mentioned criteria:
  - Provision and installation of horizontal medical bed head unit designed to meet the needs of general hospitals' wards.
  - Installation, fixation and testing of medical gas piping
  - Provision of an essential gas outlet control device for the medical gas pipeline system,
  - Equipped with outlet for Oxygen, suction and compressed air
  - Gas terminal cover to keep the gas terminal clean and hygienic
  - Each bed head unit includes its own medical gas and oxygen regulator, which must be resistant to impact.
  - The gas pipeline should be tested for 100% air tightness
  - The color of bed head unit with panels and side panels should be white
  - The bed head unit should be made of high strength industrial grade aluminum alloy material, through the whole electrostatic spraying treatment and wear-resistant.
  - Completely modular and adaptable to the needs
  - The length should be 1.40 meter
  - Gas outlet according to the hospital standards
  - LED light in both top and bottom
  - 6 power sockets accept different standards and voltages
  - Continuous Double side lamps
  - Each bed head unit includes two aluminum rails for holding the patient's information and data sheet.
  - Designed in accordance with HTM 2022 standards
  - Delivery Date to be within one month after awarding the tender



- Installation Date to be done when the location is ready after an official announcement from Heevie
- 4- The materials involved within the tender must be of good origin and have a quality certificate showing that.
- 5- The material shall be in accordance with the specifications in the tender offer list and otherwise, the tender shall be canceled. If any material is found to be in violation of the specifications mentioned upon receipt from the Company/Supplier that will be awarded the tender, Heevie shall have the right to cancel the agreement.
- 6- In case of delay in preparing and processing in the agreed period, the Company/Supplier shall bear a delay penalty of (150,000) one hundred fifty thousand Iraqi dinars for each day of delay, and in case of delay more than one month, the contract shall be canceled.
- 7- The provided tender must be completed regarding the single and the total accounts, otherwise it will be neglected.
- 8- The Company/Supplier for which the tender is awarded should have been taxed with the tax authority.
- 9- The Company/Supplier shall bear full responsibility in the event of any changes in the original materials or manufacturing defects, and it will have to be retrieved or replaced.
- 10- The Company/Supplier undertakes that the prices of the materials shall be within the normal rate of prices. Otherwise, it shall bear all the financial and legal implications.
- 11- The Company/Supplier for which the tender is awarded shall bear the transport and transfer fees to the destination.

